



WHEN PARTICLE SIZE MATTERS

Operating instructions Octagon 200





Technical specifications

Machine type designation: Octagon 200

General Information

The sieve shaker Octagon 200 is suitable for all sieving tasks in laboratories as well as onsite and provides optimum sieving action for fast and reproducible results.

It is robust, compact and sufficiently lightweight to be portable. Its electromagnetic drive combined with a 3D sieving motion ensures excellent separation efficiency in a short amount of time.

A digital display as well as a quick-release clamping system makes operation very easy and straightforward.

Advantages

- Easy-to-use sieve clamping system
- Accepts up to 8 full height 200 mm & 8" diameter sieves
- Suitable for dry and wet sieving
- 10 amplitude settings & digital timer
- 3D sieving motion allows for high separation efficiency and non-blinding sieving action
- Different voltages available
- No mechanical moving parts
- Compact & portable



Do not make any changes to the machine and use only spare parts and accessories approved by Endecotts Ltd.

The declaration of conformity to the European directives by Endecotts will otherwise lose its validity.

Furthermore this will result in the loss of any kind of guarantee claims.



WHEN PARTICLE SIZE MATTERS

Specifications

Range	20 µm to 125 mm
Drive / sieving motion	electromagnetic 3D
Max. Batch / feed capacity	3 kg
Max. Number of fractions	8 full height / 16 half height (200mm sieves)
Amplitude	0 - 3 mm digital setting in 10 steps
Speed	3,000 min ⁻¹ at 50 Hz
Time display	digital, 0:10-99:59 min
Interval operation	yes (one mode)
Suitable for dry sieving	yes
Suitable for wet sieving	yes
Serial interface	-
Sieve diameter	100 / 200 mm / 8"
Max. Height of sieve stack	up to 450 mm
Clamping device	quick-release clamping system (included)
Model	bench top
Protection code	IP 54
Electrical supply	different voltages available
Power connection	1 – Phase
W x H x D	418 x 232 x 435 mm
Net weight	43 kg
Standards	CE

Noise characteristic values:

Example 1:

Emission value related to workplace $L_{pAeq} = 63$ dB (A)

(Operating conditions: Material to be sieved = quartz sand, grain size < 1mm, 5 sieves Amplitude = 1.5 mm)

Example 2:


Emission value related to workplace $L_{pAeq} = 67$ dB (A)


(Operating conditions: Material to be sieved = quartz sand, grain size < 1mm, 5 sieves Amplitude = 3 mm)

Controls and functions

The sieve shaker should be placed on a level surface to ensure symmetrical distribution of the sample over the sieve mesh. The surface should be rigid and robust where vibration will not cause problems.

Operators should be familiar with, and fully understand the controls and indicators before operating the machine. This should be done in conjunction with the diagram below and control panel description.





1	LED light to indicate interval operation ON.
2	Interval mode button M1 switches interval operation ON/OFF, upper LED lights up; 10 sec's on, 2 sec off
3	Display shows the preselected sieving time, 00:10 – 99:59 sec's.
4	LED light to indicate START button ">" ON
5	Start the machine by pressing the START button ">".
6 & 8	"<" & ">" button reduces/increases the sieving time, 00:10 – 99:59 sec's.
7	Stop the machine by pressing the STOP button "□".
9 & 10	"<" & ">" button reduces/increases the amplitude, in 10 steps
11	10 LED's for the Power bar indicating amplitude setting in 10 steps

Setting Up

Electrical connections

- Ensure that the voltage and frequency on the Rating Label, at the rear of the shaker correspond with the local electrical mains supply.
- Connect the Octagon 200 to the power supply using the connection cable provided.



Do not connect to any other supply other than stated on the nameplate, otherwise electrical and mechanical components can be damaged.

Ambient temperature: 5°C to 40°C



If the ambient temperature is exceeded or drops below the specified value the electrical and mechanical components may become damaged and the performance data can change to an unknown degree.


Atmospheric humidity:

Maximum relative humidity 80% at temperatures up to 31°C, with linear reduction down to 50% relative humidity at 40°C.



At high atmospheric humidity the electrical and mechanical components may become damaged and the performance data can change to an unknown degree.

Transport protection



Place the **Octagon 200** on a stable laboratory bench, since otherwise unpleasant vibrations will be transmitted. Being placed on a level surface ensures symmetrical distribution of the sample over the sieves, during operation.

- Unscrew the two hexagon screws on the bottom of the sieve shaker with an open-jawed spanner until the sieve plate is movable.
- Keep hexagon screws (transport protection) for possible transport at a later date.





If the machine is operated with transport protection devices, or is transported without transport protection, mechanical components may become damaged.

Reusing transport protection device

- Dismount the clamping unit.
- Screw the hexagon screws into the two holes on the bottom of the sieve shaker and tighten with open-jawed spanner until the sieve plate is no longer movable.

Mount sieve clamping unit

Fit one M12 nut onto each clamp rod selected for use, then screw the pair of clamp rods into the location plate and tighten the locknuts.



Place the two large plain washers over the 20mm threads of the clamp sleeves. These can be seen protruding vertically from the two side lugs on the clamp plate. Screw the two hand wheels loosely onto the 20mm threads; leave a gap of 5 mm between the large washer and the hand wheel face. Do not tighten right down at this point.


Electrical connection

- The voltage and frequency for the **Octagon 200** is given on the type plate.
- Ensure that the values agree with the existing power supply.
- Connect the **Octagon 200** to the power supply using the connection cable provided.
- When connecting the power cable to the mains external fusing is necessary according to the regulations of the installation location.




If the values given on the type plate are not observed electrical and mechanical components can be damaged.

Operation Instructions



The sieve set is made up of the collecting bottom and the analytical sieves, which are arranged one on top of the other according to increasing mesh size and a the final lid.

In order to guarantee exact results under fast sieving conditions, the quantity of material to be sieved should be adapted to the sieve diameter and the nominal mesh size. More detailed information is displayed in our "TEST SIEVING MANUAL".





WHEN PARTICLE SIZE MATTERS

Place the receiver centrally on the location casting. Stack the required sieves on top of the receiver. Put the samples in the top sieve and fit the lid.

Align the locking assemblies in figure 1 the two side lugs of the clamp plate with the round clamp rods. Slide the clamp plate down squarely onto the lid at the top of the sieve stack. Ensure that the clamping hand wheels at the top (figure 2) are loose and the locking assemblies are fully pushed down. There should be a 5mm gap between the large plain washer and the face of the hand wheel. Place one hand on the top of the clamp plate (figure 3) and hold square while locking one side handle lever. Repeat for the opposite side handle lever. The side handle levers can be set vertically downwards by pressing on the Red button and pulling the handle outwards to release (figure 4). Turn the handle to a safe, convenient angle downwards and release to engage the teeth. Screw the two clamping hand wheels down simultaneously to ensure the clamping plate is square. Continue until the hand wheels are tight against the internal stop. Hand tightness must be exerted so that the assembly does not loosen during vibration. The locking side handle levers and clamping hand wheels must be tightened sufficiently to ensure that the sieves and receiver are clamped securely during operation.



Damage may occur if the shaker is allowed to operate with a loose clamping plate.

Wet sieving

The Wet Sieving Adaptor Kit is supplied as an optional extra for 8" or 200mm diameter sieves and should be ordered separately.

The Wet Sieving Adaptor Kit consists of the following items:

- 1 Off Special Wet Sieving Clamp Plate.
- 1 Set 'O' Ring Seals. One required for each sieve in the stack.
- 1 Off Special Wet Sieving Receiver.
(Specify for 200mm or 8 inch diameter sieves)
- 1 Off Nylon Hose tail.



The nylon hose tail must be fitted. The hose tail must have a suitable length of hose fitted to drain into a convenient drainage point. The 'O' ring seals are fitted on to the outside of the bottom rim of each sieve, so that when the sieves are stacked onto each other they form a seal. The bottom sieve is stacked on the special receiver with a spout. Fit a suitable length of hose to the spout to drain into a convenient drainage point. The clamp plate is usually supplied with the rose reversed to avoid damage. Undo and reverse, so that the rose head is on the inside. Remove the lid from the sieve stack and replace the standard clamp plate with the wet sieving clamp plate. Fit a suitable length of hose to the inlet of the rose on the clamp plate and connect to the fluid supply with flow regulation.



Never operate your **Sieve shaker** directly in water.
Danger through current surge.



During wet sieving always operate your **Sieve shaker** connected to a mains socket protected by an FI protective (safety) switch.




The water quantity added should always be dosed in such a way that the sieve surface is only just wetted.



Use vent rings

Accumulation of water in the sieve set can cause overloading and therefore damage or destruction of the wire mesh. Vent rings are necessary in order to prevent water accumulation.



**CERTIFICATE OF CE-CONFORMITY
ANALYTICAL SIEVE SHAKER
OCTAGON 200**

Certificate of CE-Conformity according to:

EC Mechanical Engineering Directive 2006/42/EC

Applied harmonized standards, in particular:
EN ISO 12100 Security of machines

EC Directive Electromagnetic Compatibility 2004/108/EC

Applied standards, in particular:

EN55011:2009+A1:2010, Group 1, Class B Radio disturbance characteristics –
Limits and methods of measurement

EN 61000-3-2:2006+A1:2009+A2:2009
EN 61000-3-3:2008
EN61326-1:2006

Additional applied standards, in particular

EN 61010 Safety prescriptions concerning measuring, operating, controlling and laboratory equipment

Authorised for the compilation of technical documents:

Endecotts Ltd (technical documentation)

The following records are held by Endecotts Ltd in the form of Technical Documentation:

Detailed records of engineering development, construction plans, study (analysis) of the measures required for conformity assurance, analysis of the residual risks involved and operating instructions in due form according to the approved regulations for preparation of user information data.

The CE-conformity of the Endecotts Analytical Sieve Shaker Type Octagon 200 is assured herewith.

In case of a modification to the machine not previously agreed with us as well as the use of not licensed spare parts and accessories this certificate will lose its validity.

Endecotts Ltd

London, July 2014

