

Certificate of Analysis

IMMUNOSCAN CCPlus®

REF RA-96 PLUS

LOT ZS 1885 2020-10-01

Reagents			LOT		Specification	Result
Ag			ZS 1519	2021-04-01		
CAL	A		ZS 1956	2020-12-25	OD	1.82
CAL	B		ZS 1957	2020-12-25	OD	1.56
CAL	C		ZS 1958	2020-12-25	OD	0.99
CAL	D		ZS 1959	2020-12-25	OD	0.39
CAL	E		ZS 1960	2020-12-25	OD ≥0.15	0.21
CONTROL	REF		ZS 1961	2020-12-25	OD ≥0.15	0.22
CONTROL	+		ZS 1963	2020-12-25	180-340 U/mL	254
					Ratio 4.2-7.6	5.1
CONTROL	-		ZS 1962	2021-06-24	OD	0.01
					Ratio <0.95	0.04
DIL			ZS 1967	2021-06-23		
CONJ			ZS 1968	2020-12-29		
BUF	WASH	20X	ZS 1965	2021-06-23		
SUBS		TMB	ZS 1305	2020-10-01		
H ₂ SO ₄		0.5M	ZS 1955	2021-06-23		

It is hereby certified that this LOT has been manufactured, tested and approved according to our quality procedures.

Date of approval: 2019-07-25

Gordana Pajic Peka

Quality Approved


Svar Life Science AB
Postal address

 P.O. Box 50117
 SE-202 11 Malmö
 Sweden

Visiting address

 Lundavägen 151
 SE-212 24 Malmö
 Sweden

www.svarlifescience.com