

Certificate of Analysis

IMMUNOSCAN CCPlus®

REF RA-96 PLUS

LOT ZS 1952 2020-07-08

Reagents			LOT		Specification	Result
Ag			XS 2673	2020-11-30		
CAL	A		ZS 1142	2020-07-29	OD	2.27
CAL	B		ZS 1143	2020-07-29	OD	1.93
CAL	C		ZS 1144	2020-07-29	OD	1.13
CAL	D		ZS 1145	2020-07-29	OD	0.43
CAL	E		ZS 1146	2020-07-29	OD ≥0.15	0.23
CONTROL	REF		ZS 1147	2020-07-29	OD ≥0.15	0.23
CONTROL	+		ZS 1148	2020-07-29	180-340 U/mL	262
					Ratio 4.2-7.6	5.7
CONTROL	-		ZS 1149	2021-01-18	OD	0.02
					Ratio <0.95	0.10
DIL			ZS 1151	2021-01-18	OD	0.06
CONJ			ZS 1150	2020-07-29		
BUF	WASH	20X	ZS 1152	2021-02-01		
SUBS		TMB	XS 2719	2020-08-31		
H ₂ SO ₄		0.5M	ZS 1153	2021-01-26		

It is hereby certified that this LOT has been manufactured, tested and approved according to our quality procedures.

Date of approval: 2019-06-25

Quality Approved



Svar Life Science AB

Postal address
 P.O. Box 50117
 SE-202 11 Malmö
 Sweden

Visiting address
 Lundavägen 151
 SE-212 24 Malmö
 Sweden



www.svarlifescience.com