

## CERTIFICATE OF ANALYSIS

**32X LB13 (batch D)**

### Certified Reference Material Information

Type: LEADED BRONZE (CHILL CAST)  
Form and Size: Disc ~40mm diameter  
Manufactured by: Maybrey Reliance Foundry  
Certified and Supplied by: MBH Analytical Ltd

### Assigned Values

#### Percentage element by weight

Element	Sn	Pb	Zn	Fe	Ni	Si	P	Co
Value <sup>1</sup>	5.98	7.04	0.77	0.0159	0.629	0.008	0.033	0.0044
Uncertainty <sup>2</sup>	0.02	0.04	0.01	0.0006	0.006	0.001	0.001	0.0002

Element	As	Bi	Sb	Ag	S	Al	Cd	Cu
Value <sup>1</sup>	0.118	0.058	0.092	0.0249	0.074	0.0014	0.0006	85.1
Uncertainty <sup>2</sup>	0.005	0.002	0.002	0.0009	0.004	0.0004	0.0002	0.1

Note: values given in parentheses are not certified - they are provided for information only.

### Definitions

- The certified values are the present best estimates of the true content for each element. Each value is a panel consensus, based on the averaged results of an interlaboratory testing programme, detailed on page 3.
- The uncertainties are value judgements, based on the 95% confidence interval derived from the wet analysis results, in combination with a statistical assessment of the homogeneity data, as described on page 2.

### Certified by:

MBH ANALYTICAL LIMITED



on 12<sup>th</sup> March 2020

C Eveleigh

## **Method of Preparation**

This reference material was produced from commercial-purity metals, and master alloys. The discs are the product of one melt poured into a sequence of multiple chill moulds with feeding systems designed to ensure sound discs. Approximately 2mm has been removed from the cast faces of the discs to minimise surface effects.

## **Sampling**

Samples for chemical analysis were taken from various positions throughout the casting process. At least 15% of all discs were selected for non-destructive homogeneity testing.

## **Homogeneity**

Samples representative of the batch were checked for uniformity using an optical emission spectrometer.

For all accepted material, through-batch variation values were derived for each element as an indicator of any minor compositional variation (as determined for the specific sample size and other limitations of the spectrometer).

## **Chemical Analysis**

Analysis was carried out on millings taken from samples representative of the product. It was performed by a panel of laboratories mostly operating within the terms of EN ISO/IEC 17025, using documented standard reference methods and validated by appropriate reference materials.

The individual values listed overpage are the average of each analyst's results.

## **Estimation of Uncertainties**

Each element certified has been analysed by several laboratories, and 95% half-width confidence intervals ( $C_{(95\%)}$ ) for the resultant mean values have been derived by the method shown on page 3.

As a separate exercise, the degree of non-homogeneity of the batch for each element has been quantified by a programme of non-destructive application testing, described above.

The final certified uncertainty for each element has been derived by combining these two factors, using the square-root of the summed squares.

## **Traceability**

Much of the analytical work performed to assess this material has been carried out by laboratories with proven competence, as indicated by their accreditation to ISO 17025. It is an implicit requirement for this accreditation that analytical work should be performed with due traceability, via an unbroken chain of comparisons, each with stated uncertainty, to primary standards such as the mole, or to nationally- or internationally-recognised reference materials.

Of the individual results herein, some have traceability (to the mole) via primary analytical methods. Some are traceable to substances of known stoichiometry. Most have traceability via commercial solutions. Furthermore, some results have additional traceability to NIST standards, as part of the analytical calibration or process control.

## **Usage**

Intended use: With optical emission and X-ray fluorescence spectrometers.

Recommended method of use: Copper alloys are generally prepared by machining on a mill or a lathe. However, users are recommended to follow the calibration and sample preparation procedures specified by the relevant instrument manufacturer.

Preparation should be the same for reference materials and the samples for test.

A minimum of five consistent replicate analyses is recommended to provide the necessary sample size. Users are advised to check against possible bias between reference materials and production samples due to differences in metallurgical history, and be aware of possible inter-element effects.

## Analytical Data

### Percentage element by weight

Sample	Sn	Pb	Zn	Fe	Ni	Si	P	Co
1	5.930	6.954	0.7445	0.0137	0.6130	0.0063	0.0302	0.0038
2	5.956	6.969	0.7456	0.0144	0.6143	0.0069	0.0306	0.0040
3	5.980	6.983	0.7485	0.0149	0.6190	0.0075	0.0319	0.0041
4	5.984	6.986	0.7512	0.0155	0.6220	0.0077	0.0325	0.0042
5	5.990	7.015	0.7560	0.0155	0.6280	0.0079	0.0333	0.0043
6	5.993	7.024	0.7645	0.0157	0.6307	0.0082	0.0340	0.0043
7	5.999	7.040	0.7680	0.0158	0.6323	0.0091	0.0342	0.0045
8	6.005	7.050	0.7766	0.0160	0.6330	0.0092	0.0344	0.0045
9	6.012	7.079	0.7778	0.0160	0.6373	0.0103	0.0355	0.0046
10		7.094	0.7790	0.0162	0.6392	0.0104	0.0356	0.0046
11		7.154	0.7882	0.0167	0.6414		0.0360	0.0049
12		7.175	0.7903	0.0168	0.6416			0.0049
13			0.7970	0.0174				
14				0.0175				
<b>Mean</b>	<b>5.983</b>	<b>7.044</b>	<b>0.7682</b>	<b>0.0159</b>	<b>0.6293</b>	<b>0.0084</b>	<b>0.0335</b>	<b>0.0044</b>
<b>Std Dev</b>	0.026	0.071	0.0180	0.0011	0.0102	0.0014	0.0020	0.0003
<b>C<sub>(95%)</sub></b>	0.02	0.04	0.01	0.0006	0.006	0.001	0.001	0.0002

Sample	As	Bi	Sb	Ag	S	Al	Cd	Cu
1	0.1110	0.0550	0.0868	0.0222	0.0698	0.0008	0.0004	84.88
2	0.1113	0.0553	0.0874	0.0242	0.0702	0.0008	0.0004	85.01
3	0.1114	0.0555	0.0897	0.0245	0.0704	0.0010	0.0004	85.09
4	0.1164	0.0577	0.0902	0.0247	0.0708	0.0014	0.0004	85.13
5	0.1197	0.0579	0.0920	0.0250	0.0756	0.0014	0.0004	85.19
6	0.1205	0.0581	0.0920	0.0250	0.0759	0.0018	0.0007	85.19
7	0.1240	0.0584	0.0924	0.0251	0.0784	0.0018	0.0009	85.23
8	0.1250	0.0605	0.0931	0.0254	0.0820	0.0018	0.0010	85.33
9	0.1251	0.0608	0.0941	0.0264				
10			0.0952	0.0268				
11			0.0957					
<b>Mean</b>	<b>0.1183</b>	<b>0.0577</b>	<b>0.0917</b>	<b>0.0249</b>	<b>0.0741</b>	<b>0.0014</b>	<b>0.0006</b>	<b>85.13</b>
<b>Std Dev</b>	0.0060	0.0021	0.0029	0.0013	0.0045	0.0004	0.0003	0.14
<b>C<sub>(95%)</sub></b>	0.005	0.002	0.002	0.0009	0.004	0.0004	0.0002	0.1

Note: C<sub>(95%)</sub> is the 95% half-width confidence interval derived from the equation:

$$C_{(95\%)} = (t \times SD) / \sqrt{n}$$

where n is the number of available values, t is the Student's t value for n-1 degrees of freedom, and SD is the standard deviation of the test results.

## Participating Laboratories

Element Ltd  
Sheffield Assay Office  
Anchorcert Analytical  
Universal Scientific Laboratory Pty Ltd  
Genitest, Inc  
Shanghai Jinyi Test Technology Co  
Luo Yang Copper  
Raghavendra Spectromet Laboratory  
TCR Engineering Services Pvt Ltd  
Institute of Non-Ferrous Metals  
INCDMNR-IMNR  
Mineral & Metallurgical Laboratories  
AMG Superalloys UK Ltd  
Analyticka Laborator Lithea sro

Middlesbrough, England  
Sheffield, England  
Birmingham, England  
Milperra, NSW, Australia  
Montreal, QC, Canada  
Shanghai, China  
Luo Yang, He Nan, China  
Bangalore, India  
Mumbai, India  
Gliwice, Poland  
Pantelimon, Romania  
Bangalore, India  
Rotherham, England  
Brno, Czech Republic

UKAS accreditation 0239  
UKAS accreditation 0012  
UKAS accreditation 0667  
NATA accreditation 0492  
PJ accreditation L17-153  
CNAL accreditation 0783  
CNAL accreditation 0173  
NABL accreditation 0371  
NABL Accreditation 0367  
PCA accreditation AB274

Note: to achieve the above accreditation (UKAS, etc), test houses must demonstrate conformity to the general requirements of EN ISO/IEC 17025.

## Analytical Methods Used

ELEMENT	RESULT No. & METHOD		
	ICP-AES	FAAS	OTHER
Tin	2, 5, 8	-	1, 3, 4, 7, 9 6 volumetric (iodate) photometric (phenyl fluorone)
Lead	1-3, 6, 7, 9-11	4, 8	5, 12 volumetric (sulfocyanate)
Zinc	1, 2, 4-6, 8-10, 12	3, 7, 11, 13	
Iron	1, 2, 4-8, 10, 13, 14	3, 9, 12	11 photometric (orthophenanthroline)
Nickel	2, 3, 6, 7, 9, 10, 12	4, 5, 11	1 gravimetric (dimethyl glyoxime)
Silicon	1, 2, 4-6, 10	-	8 photometric (dimethyl glyoxime)
Phosphorus	1-3, 5, 6, 8-10	-	3, 7-9 photometric (molybdenum yellow)
Cobalt	1, 3-6, 8, 10, 11	2, 7, 9, 12	4, 7, 11 photometric (molybdenum yellow)
Arsenic	1-5, 7-9	6	
Bismuth	1, 3-8	2, 9	
Antimony	1, 3, 5-9, 11	2, 10	4 photometric (crystal violet)
Silver	3, 4, 6-10	1, 2, 5	
Sulfur	3, 6	-	1, 2, 4, 5, 7, 8 combustion (IR or volumetric detection)
Aluminium	4-8	1, 3	2 photometric (chrome azurol S)
Cadmium	1-4, 7, 8	5, 6	
Copper	1, 4	-	2, 3, 5, 7 electrogravimetric 6, 8 volumetric (thiosulfate)

## Notes

This Certified Reference Material has been produced and certified, wherever possible, in accordance with the requirements of ISO 17034 and the associated Guides, taking into account the requirements of the ISO Guide to the Expression of Uncertainty in Measurement (GUM).

The unidirectional solidification effects associated with this method of chill casting, have led to the formation of inhomogeneous segregates in the rear portion of the disc. The above certification is therefore only applicable from the front face of the disc, to a depth of 10mm. Material to the rear of the disc, to a depth of ~5 mm, is not certified.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture. All production records will be retained for a period of 20 years from the date of this certificate. Technical support for this certification will therefore expire in November 2039, although we reserve the right to make changes as issue revisions, in the intervening period.

This material is also available in the form of chippings, for the calibration or control of methods involving dissolution.

The manufacture, analysis and certification of this product were supervised by C Eveleigh, PhD, Technical Director, MBH Analytical Ltd.

The material to which this certificate of analysis refers is supplied subject to our general conditions of sale.