


3.4 Minimum Floor Space and Clearance Requirements


Model 6302G & 6302GADH

6302G OR 6302GADH SUPINE TUB
1200 TRAVERSE STRETCHER LIFT



4 INSTALLATION

4.2 Tub Dimensions and Connection Details



Connection	Location	Type
C	10" (254 mm) from centerline, 5.5" (140 mm) up from floor	3/4" Cold water stub (with recommended shut off valve)
D	5" (127 mm) from centerline, 29" (736 mm) from tub back edge	2" Drain (floor drain recommended)
E	10.5" (267 mm) from centerline	115V AC, 60 Hz 15 A Input
H	6.5" (165 mm) from centerline, 5.5" (140 mm) up from floor	3/4" Hot water stub (with recommended shut off valve)