

QUALITY CONTROL CERTIFICATE

REDIPOR PREPARED MEDIA



Product Code : 100130
 Description Tryptone Soya Agar
 90mm Petri Dishes

 Batch Number : 15061533
 Date of Manufacture: 15 June 2015
 Expiry Date : 10 August 2015

 pH (T= 7.3 ± 0.2) 7.4
 Sterility Test: PASS

Positive Control organisms to show a ratio of 0.7 - 2.0 (Inoculum level 10-100 cfu)

Organism	Reference	Productivity Ratio
Staphylococcus aureus	ATCC 6538	0.9
Staphylococcus epidermidis	ATCC 12228	1.0
Escherichia coli	ATCC 8739	1.1
Pseudomonas aeruginosa	ATCC 9027	1.2
Bacillus subtilis	ATCC 6633	1.1
Candida albicans	ATCC 10231	0.9
Aspergillus brasiliensis	ATCC 16404	1.0

Comments:

Signed By: *J. Deben*

Authorised Signatory

Date: 22 June 2015

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C