

# Ventyv<sup>®</sup>

## Nitrile PF Exam

### SELECT 3.0 Cobalt Blue



Ventyv<sup>®</sup>  
Outsmart Infection<sup>™</sup>



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |  
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607  
Issued: 20210913  
R2

# Ventyv®

## Nitrile PF Exam

### SELECT 3.0 Cobalt Blue



Ventyv® Nitrile Powder-Free Select 3.0 Cobalt Blue (Walrus) Exam Glove has a natural fit that allows for unrestricted hand motion and reduces hand fatigue. Don't let the name fool you, it's the lightest glove of the bunch.



## Walrus



Reorder #	Size	Gloves/ Box <sup>1</sup>	Boxes/ Case
10345102	XS	300	10
10346102	S	300	10
10347102	M	300	10
10348102	L	300	10
10349102	XL	250	10

### PRODUCT ATTRIBUTES

- Cobalt Blue
- Nitrile<sup>2</sup>
- Powder-free
- Polymer Coated
- Tested for Use with Chemotherapy Drugs<sup>4</sup>
- Fingertip Textured

### PHYSICAL ATTRIBUTES<sup>3</sup>

- Length: 220 mm (XS, S)  
230 mm (M, L, XL)
- Width: 70 ± 10 mm (XS)  
80 ± 10 mm (S)  
95 ± 10 mm (M)  
110 ± 10 mm (L)  
120 ± 10 mm (XL)

### THICKNESS (Single Wall)

- Finger: 0.08 mm / 3 mils
- Palm: 0.06 mm / 2 mils

### PHYSICAL PROPERTIES

- Tensile strength before aging: 27 MPa
- Tensile strength after aging: 25 MPa
- Elongation before aging: 522%
- Elongation after aging: 487%



80%

Inner boxes made with 80% recycled material

90%

Outer cartons made with 90% recycled material

1. By weight

2. This product is not made with Natural Rubber Latex

3. Physical Attributes meet or exceed ASTM standards for exam gloves

4. CAUTION: Gloves used for protection against chemotherapy drug exposure must be selected specifically for the type of chemicals used.

Review material safety data sheets for the drug being used to determine the required level of protection. Contact Ventyv for specific testing information.

Ventyv®  
Outsmart Infection®



sritrangusa.com | ventyv.com | Hello@sritrangusa.com | 1.888.483.6898 |  
+1 (813) 606-4301 | 5820 West Cypress Street, Suite H | Tampa, FL 33607  
Issued: 20210913

R2