

CERTIFICATE OF ANALYSIS

Page 1 of 1

CUSTOMER
 BioCote Ltd
 Wolverhampton Science Park
 Technology Centre
 Glaisher Drive
 Wolverhampton
 WV10 9RU.

CERTIFICATE NO. 1013404.91/4433
 CUSTOMER REF. 10/080

SAMPLE DETAILS DATE RECEIVED 01/03/2010
 CONTOUR CASINGS ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

DATE ANALYSED 03/03/2010 DATE REPORTED 05/03/2010

RESULTS (AS CFU CM²)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (INITIAL)	
		0 hrs	24 hrs	Log 10	%
29.B4000.102 WHITE (4700110) POLYSTYRENE	<i>E coli</i>	2.2E+04	1.7E+01	3.1	99.92%
29.B4000.102 WHITE (4700110) POLYSTYRENE	MRSA	2.5E+04	< 11.11	≥ 3.35	≥ 99.96%
	MRSA	2.5E+04	8.0E+02		



Key: NS = Poor survival on control supplied.

The above data shows the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the initial population.

Certificate valid for up to 2 years from date of test.

IMSL MICROBIOLOGICAL SERVICES LTD
 PALE LANE
 HARTLEY WINTNEY
 HANTS RG27 8DH
 UK

MANAGING DIRECTOR
 Peter D Askew